

"CORRIGENDUM"

E-Tender No. PPDCM/4(9)/2016-17/GC-MS DATED-15.11.2016
Dated: 02.12.2016

Reference e-Tender Notice PPDCM/4(9)/2016-17/GC-MS DATED-15.11.2016, issued vide this office of undersigned on 02.12.2016; the following changes in the e-Tender Notice are hereby notified through this Corrigendum for all concerned Suppliers.

1. At Serial no. iv of the page 3 mentioning the technical specifications of Item, in the clause regarding **Mass spectrophotometer** in the tender document:
"EI source with maximum temperature of 350 degree C or better"
'may now be read as'
"EI source with maximum temperature of 300° C or better"
2. At Serial no. v of the page 3 mentioning the technical specifications of Item, in the clause regarding **Mass Range** in the tender document:
"Upto 10000 amu/sec or better"
'May now be read as'
"Upto 1000 amu/sec or better "
3. At Serial no. vi of the page 3 mentioning the technical specifications of Item, in the clause regarding **Scan rate** in the tender document:
"Scan rate of 12000 amu/sec"
'May now be read as'
"Scan rate of 10000 amu/sec or better."



Sunil Gupta
Principal Director